



VERIFICATION OF COMPLIANCE

This Verification of Compliance is hereby issued to the below named company. The test results of this report relate only to the tested sample identified in this report.

**Technical Standard: EMC DIRECTIVE 2014/30/EU
(EN 61000-6-4 / EN 61000-6-2)**

General Information

Applicant: AXIOMTEK CO., LTD.
8F., No.4, Lane 235, Baoqiao Road, Xindian District,
New Taipei City 231, Taiwan (R.O.C.)

Product Description

EUT Description: Industrial Fieldbus
Brand Name: AXIOMTEK
Model Number: IFB122

Measurement Standard

EN 61000-6-4: 2007 + A1: 2011

EN 61000-3-2: 2014

EN 61000-3-3: 2013

EN 61000-6-2: 2005 / AC: 2005

(IEC 61000-4-2: 2008; IEC 61000-4-3: 2006 + A1: 2007 + A2: 2010; IEC 61000-4-4: 2012;
IEC 61000-4-5: 2014; IEC 61000-4-6: 2013; IEC 61000-4-8: 2009; IEC 61000-4-11: 2004)

Measurement Facilities

Xindian Lab.: **Compliance Certification Services Inc.**
No.163-1, Zhongsheng Rd., Xindian Dist., New Taipei City, 23151 Taiwan.
Tel: +886-2-22170894 / Fax: +886-2-22171029

This device has been shown to be in compliance with and was tested in accordance with the measurement procedures specified in the Standards & Specifications listed above and as indicated in the measurement report number: T161122D13-E

Sam Hu / Assistant Manager

Date: November 29, 2016

程智科技股份有限公司
Compliance Certification Services Inc.