



## VERIFICATION OF COMPLIANCE

*This Verification of Compliance is hereby issued to the below named company. The test results of this report relate only to the tested sample identified in this report.*

**Technical Standard: EMC DIRECTIVE 93/42/EEC  
(EN 60601-1-2)**

### General Information

Applicant: AXIOMTEK CO., LTD.  
8F., No.4, Lane 235, Baoqiao Road, Xindian District,  
New Taipei City 231, Taiwan (R.O.C.)

### Product Description

EUT Description: Medical Panel PC  
Brand Name: AXIOMTEK  
Model Number: MPC175-873; MPC225-873

### Measurement Standard

IEC 60601-1-2: 2007; EN 60601-1-2: 2007  
CISPR 11: 2009 + A1: 2010; EN 55011: 2009 + A1: 2010 (Group 1, Class B)  
IEC 61000-3-2: 2005 + A1: 2008 + A2: 2009; EN 61000-3-2: 2006 + A1: 2009 + A2: 2009  
IEC 61000-3-3: 2013; EN 61000-3-3: 2013  
IEC 61000-4-2: 2008; EN 61000-4-2: 2009; IEC 61000-4-3: 2006 + A1: 2007 + A2: 2010;  
EN 61000-4-3: 2006 + A1: 2008 + A2: 2010; IEC 61000-4-4: 2012; EN 61000-4-4: 2012;  
IEC 61000-4-5: 2005; EN 61000-4-5: 2006; IEC 61000-4-6: 2008; EN 61000-4-6: 2009; IEC 61000-4-8: 2009;  
EN 61000-4-8: 2010; IEC 61000-4-11: 2004; EN 61000-4-11: 2004

### Measurement Facilities

Xindian Lab.: **Compliance Certification Services Inc.**  
No.163-1, Jhongsheng Rd., Xindian Dist., New Taipei City, 23151 Taiwan.  
Tel: +886-2-22170894 / Fax: +886-2-22171029

*This device has been shown to be in compliance with and was tested in accordance with the measurement procedures specified in the Standards & Specifications listed above and as indicated in the measurement report number: T140416D14-E*

Sam Hu / Assistant Manager

Date: April 23, 2014



程智科技股份有限公司  
Compliance Certification Services Inc.